

## **EC DECLARATION OF CONFORMITY**

#### Manufacturer:

#### **ELDO CARE CORPORATION LIMITED**

Rm.1902, Easey comm. Bldg., 253-261 Hennessy Road, Wanchai, Hongkong

# **European Authorized Representative**

Company: Lotus NL B.V.

Address: Koningin Julianaplein 10, 1e Verd, 2595AA, The Hague, Netherlands.

#### Declares that the MDD described hereafter

#### Product name and model:

ELFM3600 - Non-Woven Face Mask, Ear-loop, 3-ply, non-sterile

Meet the provisions of the Council Directive 93/42/EEC & 2007/47/CE which apply to them.

Harmonized standards or normative documents: EN 14683:2019+AC:2019 TYPE IIR BFE ≥98%

The medical device has been assigned to class I according to Annex VII of the Directive 2007/47/CE. It bears the mark:

CONFORMITY ASSESSMENT ROUTE: Annex VII

This Declaration of conformity is valid in connection with the release document for the respective batch of produced devices.

The above mentioned declaration of conformity is exclusively under the responsibility of: ELDO CARE CORPORATION LIMITED

Place , date Shanghai, China Sep. 9<sup>th</sup>, 2020 Legally binding signature, Function Co-founder & CEO

**Ellen Wang** 

**EC Declaration of Conformity** 

Authorized Signature(s)

CO., LIMITED

For and on behalf of ELDO CARE



ELDO CODE	GIMA CODE	PRODUCT DESCRIPTION	PACKING
ELFM3600	20670	GISAFE – color: light blue Surgical mask 3ply - Type IIR non-woven - earloop- non sterile – singly packed	box 50pcs
ELFM3600	20700	GISAFE – color: light blue Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	box 50pcs
ELFM3600	20695	GISAFE – color: light blue Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	flowpack 10pcs
ELFM3600	20697	GISAFE – color: pink Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	box 50pcs
ELFM3600	20694	GISAFE – color: pink Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	flowpack 10pcs
ELFM3600	20698	GISAFE – color: white Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	box 50pcs
ELFM3600	20699	GISAFE – color: light green Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	box 50pcs
ELFM3600	20685	GISAFE – fantasy: cartoon Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	flowpack 10pcs
ELFM3600	20686	GISAFE – fantasy: peace Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	flowpack 10pcs
ELFM3600	20687	GISAFE – fantasy: stars Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	flowpack 10pcs
ELFM3600	20680	GISAFE – color: light blue Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	flowpack 10pcs
ELFM3600	20682	GISAFE – fantasy: cartoon Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	flowpack 10pcs
ELFM3600	20696	GISAFE – color: light blue Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	box 2000pcs (200 flowpack of 10 pcs)
ELFM3600	20701	GISAFE – color: black Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	box 50pcs
ELFM3600	20688	GISAFE - fantasy: colours Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	flowpack 10pcs
ELFM3600	20689	GISAFE – fantasy: macarones Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	flowpack 10pcs
ELFM3600	20690	GISAFE – fantasy: skulls Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	flowpack 10pcs
ELFM3600	20691	GISAFE – fantasy: military Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	flowpack 10pcs

Page 2/2



# GISAFE MASCHERINA CHIRURGICA FILTRANTE 98% 3 VELI tipo IIR con elastici - adulti - peace - flowpack

**Codice: 20686** 

Categoria: Mascherine

Unità di vendita: conf 10 pz.

Quantitativi minimi: 1

Dispositivo: Dispositivo medico

Classe: I

NSIS: 1946919

CND: T020601

EAN13: 8023279206869

Descrizione: MASCHERINE FILTRANTI 98% A 3 VELI - TIPO IIR con elastici

Modello per adultiFlowpack da 10Fantasia: peace

Mascherine chirurgiche leggere e confortevoli adatte per ospedali, studi dentistici e ambulatori.

- vasta gamma: mascherine adulti in 5 colori o 7 fantasie; mascherine pediatriche in 1 colore o 1 fantasia

- disponibili in scatola da 50 o flowpack da 10

- nasello modellabile da adattare alla forma anatomica del viso

- senza fibra di vetro, efficienza di filtrazione BFE =98%

materiale non tessuto, senza lattice
1° velo esterno 28 gsm non tessuto
2° velo 25 gsm filtro meltblown

- 3° velo interno 25 gsm non tessuto

Scatola multilingue: GB, FR, IT, ES, PT, DE, SE, GR, RO, Arabo.

Caratteristiche tecniche: BFE% = 98

Pressione differenziale (Pa/cm2) < 60

CfU/g < 30

Dimensioni: 17,5 x 9,5 cm (adulti)

Conforme a EN 14683:2019 + AC:2019; DDM 93/42/CEE Classe I







中国认可 国际互认 检测 TESTING CNAS L0599

Date:October 30,2020

Page 1 of 3

#### **ELDO CARE CORPORATION LIMITED**

RM.1902, EASEY COMM. BLDG., 253-261 HENNESSY ROAD, WANCHAI, HONGKONG

The following sample(s) was/were submitted and identified on behalf of the client as:

Sample Description : (A)GIMA BRANDED 20686 - 3 PLY SURGICAL MASK - DISPOSABLE -

EARLOOP - FANTASY: PEACE

Composition : (A)67% non-woven, 33% melt-blown

Sample Color : (A)FANTASY: PEACE

Style No. : ELFM3600 Lot No. : Not provided

Manufacturer : ELDO CARE CO., LIMITED

Test Performed : Selected test(s) as requested by applicant

Sample Receiving Date : Oct 19, 2020

Testing Period : Oct 21, 2020 - Oct 30, 2020

Test Result(s) : Unless otherwise stated the results shown in this test report refer only to the

sample(s) tested, for further details, please refer to the following page(s).

Signed for and on behalf of

SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd Testing Center

Helen Xuan (Authorized Signatory)





**Test Report Test Result** 

SL52045300698601TX

Date:October 30,2020

Page 2 of 3

#### EN 14683:2019+AC:2019 Medical Face Masks-Requirements and Test Methods

#### Clause 5.2 Performance Requirement

### Clause 5.2.2 Bacterial Filtration Efficiency (BFE)

(EN 14683:2019+AC:2019 Annex B)

Sample: A

Test Side Inside

Test Area Approximately 60 cm<sup>2</sup>

Flow Rate 28.3 L/min

Pre-Conditioning : Minimum of 4 hours at 21±5°C and 85±5% R.H.

: ~175mm x 165mm Dimensions of test specimen

Positive Control Average : 2450 CFU Negative Monitor Count : < 1 CFU Mean Particle Size : 3.0 ±0.3µm

Test bacteria : Staphylococcus aureus ATCC 6538

Test Item	Specimen No.	Result
Bacterial Filtration Efficiency (BFE)	1	99.9%
	2	99.9%
	3	99.9%
	4	99.9%
	5	99.9%

- 1) Performance Requirement: Type I≥95%, Type II≥98%, Type IIR ≥98%
- The number of specimens that shall be tested is minimum 5, but can be greater and shall be increased if necessary to allow for an AQL(Acceptable Quality Level) of 4%.



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <a href="http://www.sgs.com/en/Terms-and-Conditions.aspx">http://www.sgs.com/en/Terms-and-Conditions/Terms

3<sup>rd</sup> Building, No. 889, Yishan Road, Xuhui District Shanghai, China 200233 t (86-21) 61402666 f (86-21) 64958763

中国・上海・徐汇区宜山路889号3号楼 邮编: 200233

t (86-21) 61402666 f (86-21) 64958763 e sgs.china@sgs.com




SL52045300698601TX

Date:October 30,2020

Page 3 of 3





The statement of conformity in this test report is only based on measured values by the laboratory and does not take their uncertainties into consideration.

\*\*\*End of Report\*\*\*







Date:October 30,2020



中国认可 国际互认 检测 TESTING CNAS L0599

Page 1 of 3

#### Test Report

SL52045300639501TX

**ELDO CARE CORPORATION LIMITED** 

RM.1902, EASEY COMM. BLDG., 253-261 HENNESSY ROAD, WANCHAI, HONGKONG

The following sample(s) was/were submitted and identified on behalf of the client as:

Sample Description : (A)GIMA BRANDED 20686 - 3 PLY SURGICAL MASK - DISPOSABLE -

EARLOOP - FANTASY: PEACE

Composition : (A)67% non-woven, 33% melt-blown

Sample Color : (A)FANTASY: PEACE

Style No. : ELFM3600 Lot No. : Not provided

Manufacturer : ELDO CARE CO., LIMITED

Test Performed : Selected test(s) as requested by applicant

Sample Receiving Date : Oct 19, 2020


Testing Period : Oct 20, 2020 - Oct 30, 2020

Test Result(s) : Unless otherwise stated the results shown in this test report refer only to the

sample(s) tested, for further details, please refer to the following page(s).

Signed for and on behalf of

SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd Testing Center







SL52045300639501TX

Date:October 30,2020

Page 2 of 3

Test Result

#### EN 14683:2019+AC:2019 Medical Face Masks-Requirements and Test Methods

#### Clause 5.2 Performance Requirement

#### Clause 5.2.3 Breathability

(EN 14683 :2019+AC:2019 Annex C)

Sample: A

Test Side Randomly test in different location (1 around and 4 away from the centric

point) on each of the 5 masks

**Pre-Conditioning** Minimum of 4 hours at 21±5°C and 85±5% R.H.

Test Area 4.9 cm<sup>2</sup> Flow Rate 8 I/min

Specimen No.	Test Area No.	Different Pressure for each	The average value for each test
		tested area (Pa/cm²)	specimen (Pa/cm²)
	1-1	34.6	
	1-2	38.2	
1	1-3	39.3	38
	1-4	39.2	
	1-5	39.4	
	2-1	38.8	
	2-2	38.5	
2	2-3	33.7	37
	2-4	36.5	
	2-5	38.2	
	3-1	39.6	
	3-2	38.9	
3	3-3	34.6	38
	3-4	36.2	
	3-5	38.9	
4	4-1	39.1	
	4-2	37.2	
	4-3	38.4	38
	4-4	38.3	
	4-5	36.7	
5	5-1	39.4	
	5-2	38.5	
	5-3	39.0	39
	5-4	37.3	
	5-5	38.9	7

#### Remark:

- 1) Performance Requirement: Type I<40 Pa/cm², Type II<40 Pa/cm², Type IIR<60 Pa/cm²
- The number of specimens that shall be tested is minimum 5, but can be greater and shall be increased if necessary to allow for an AQL(Acceptable Quality Level) of 4%.



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <a href="http://www.sgs.com/en/Terms-and-Conditions.aspx">http://www.sgs.com/en/Terms-and-Conditions/Terms

3<sup>rd</sup> Building, No. 889, Yishan Road, Xuhui District Shanghai, China 200233 t (86-21) 61402666 f (86-21) 64958763

中国・上海・徐汇区宜山路889号3号楼 邮编: 200233

t (86-21) 61402666 f (86-21) 64958763 e sgs.china@sgs.com



SL52045300639501TX

Date:October 30,2020

Page 3 of 3





The statement of conformity in this test report is only based on measured values by the laboratory and does not take their uncertainties into consideration.

\*\*\*End of Report\*\*\*











Date:October 30,2020

Page 1 of 3

#### **ELDO CARE CORPORATION LIMITED**

RM.1902, EASEY COMM. BLDG., 253-261 HENNESSY ROAD, WANCHAI, HONGKONG

The following sample(s) was/were submitted and identified on behalf of the client as:

Sample Description : (A)GIMA BRANDED 20686 - 3 PLY SURGICAL MASK - DISPOSABLE -

**EARLOOP - FANTASY: PEACE** 

Composition : (A)67% non-woven, 33% melt-blown

Sample Color : (A)FANTASY: PEACE

Style No. : ELFM3600 Lot No. : Not provided

Manufacturer : ELDO CARE CO., LIMITED

Test Performed : Selected test(s) as requested by applicant

Sample Receiving Date : Oct 19, 2020


Testing Period : Oct 20, 2020 - Oct 30, 2020

Test Result(s) : Unless otherwise stated the results shown in this test report refer only to the

sample(s) tested, for further details, please refer to the following page(s).

Signed for and on behalf of

SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd Testing Center







Test Report Test Result SL52045300649801TX

Date:October 30,2020

Page 2 of 3

#### EN 14683:2019+AC:2019 Medical Face Masks-Requirements and Test Methods

#### Clause 5.2 Performance Requirement

#### Clause 5.2.4 Splash Resistance

(ISO 22609:2004)

Sample: A

Test Blood Pressure : 16.0kPa

Pre-Conditioning : Minimum of 4 hours at 21±5°C and 85±5% R.H.

Distance of the mask to the tip of cannula : 300±10mm

Test Specimen#	Penetration on inside surface	Conclusion	Test Specimen#	Penetration on inside surface	Conclusion
1	None Seen	Pass	17	None Seen	Pass
2	None Seen	Pass	18	None Seen	Pass
3	None Seen	Pass	19	None Seen	Pass
4	None Seen	Pass	20	None Seen	Pass
5	None Seen	Pass	21	None Seen	Pass
6	None Seen	Pass	22	None Seen	Pass
7	None Seen	Pass	23	None Seen	Pass
8	None Seen	Pass	24	None Seen	Pass
9	None Seen	Pass	25	None Seen	Pass
10	None Seen	Pass	26	None Seen	Pass
11	None Seen	Pass	27	None Seen	Pass
12	None Seen	Pass	28	None Seen	Pass
13	None Seen	Pass	29	None Seen	Pass
14	None Seen	Pass	30	None Seen	Pass
15	None Seen	Pass	31	None Seen	Pass
16	None Seen	Pass	32	None Seen	Pass
Number of Pass:		32			
Overall result:		Acceptable			

#### Remark:

- 1) Performance Requirement Type I: N/A, Type II: N/A, Type IIR: ≥16.0kPa
- 2) Test was conducted within 60s after removal from conditioning chamber.
- 3) An acceptable quality limit of 4.0% is met for a single sampling plan when 29 or more of the 32 tested specimens show pass results.



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <a href="http://www.sgs.com/en/Terms-and-Conditions.aspx">http://www.sgs.com/en/Terms-and-Conditions/Terms

3<sup>11</sup>Building,No.889,Yishan Road,Xuhui District Shanghai,China 200233 中国・上海・徐汇区宜山路889号3号楼 邮編: 200233 t (86-21) 61402666 t (86-21) 61402666

f (86-21) 64958763 f (86-21) 64958763 www.sgsgroup.com.cn e sgs.china@sgs.com




SL52045300649801TX

Date:October 30,2020

Page 3 of 3





The statement of conformity in this test report is only based on measured values by the laboratory and does not take their uncertainties into consideration.

\*\*\*End of Report\*\*\*



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <a href="http://www.sgs.com/en/Terms-and-Conditions.aspx">http://www.sgs.com/en/Terms-and-Conditions/Terms

www.sgsgroup.com.cn







中国认可 国际互认 检测 TESTING CNAS L0599

Date:November 20,2020

Page 1 of 3

#### **ELDO CARE CORPORATION LIMITED**

RM.1902, EASEY COMM. BLDG., 253-261 HENNESSY ROAD, WANCHAI, HONGKONG

The following sample(s) was/were submitted and identified on behalf of the client as:

Sample Description : (A)GIMA BRANDED 20686 - 3 PLY SURGICAL MASK - DISPOSABLE-

EARLOOP - FANTASY: PEACE

Composition : (A)67% non-woven, 33% melt-blown

Sample Color : (A)FANTASY: PEACE

Style No. : ELFM3600 Lot No. : Not provided

Manufacturer : ELDO CARE CO., LIMITED

Test Performed : Selected test(s) as requested by applicant

Sample Receiving Date : Nov 10, 2020

Testing Period : Nov 11, 2020 - Nov 20, 2020

Test Result(s) : Unless otherwise stated the results shown in this test report refer only to the

sample(s) tested, for further details, please refer to the following page(s).

Signed for and on behalf of

SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd Testing Center

Sara Guo (Account Executive)

Dongjing Liu (Authorized Signatory)



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <a href="http://www.sgs.com/en/Terms-and-Conditions.aspx">http://www.sgs.com/en/Terms-and-Conditions.aspx</a> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <a href="http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx">http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx</a>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's esponsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or, email: CN.Doccheck@sgs.com



SL52045312279201TX

Date:November 20,2020

Page 2 of 3

#### **Test Result**

#### EN 14683:2019+AC:2019 Medical Face Masks-Requirements and Test Methods

#### Clause 5.2 Performance Requirement

#### Clause 5.2.5 Microbial Cleanliness

(EN 14683:2019+AC:2019 Annex D and EN ISO 11737-1:2018)

Sample: A

Test Specimen#	Mask Weight(g)	Total Bioburden, (CFU/mask)	Total Bioburden, (CFU/g)
1#	3.61	12	3.32
2#	3.62	15	4.14
3#	3.64	9	2.47
4#	3.58	12	3.35
5#	3.62	6	1.66

Remark: Performance Requirement: Type I≤30 CFU/g, Type II≤30 CFU/g, Type IIR≤30 CFU/g



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <a href="http://www.sgs.com/en/Terms-and-Conditions.aspx">http://www.sgs.com/en/Terms-and-Conditions/Terms

3<sup>rd</sup>Building,No.889,Yishan Road,Xuhui District Shanghai,China 200233 t (86-21) 61402666 f (86-21) 64958763

中国・上海・徐汇区宜山路889号3号楼 邮编: 200233

t (86-21) 61402666

f (86-21) 64958763

e sgs.china@sgs.com



SL52045312279201TX

Date:November 20,2020

Page 3 of 3





The statement of conformity in this test report is only based on measured values by the laboratory and does not take their uncertainties into consideration.

\*\*\*End of Report\*\*\*

