

EC DECLARATION OF CONFORMITY

Manufacturer:

ELDO CARE CORPORATION LIMITED

Rm.1902, Easey comm. Bldg., 253-261 Hennessy Road, Wanchai, Hongkong

European Authorized Representative

Company: Lotus NL B.V.

Address: Koningin Julianaplein 10, 1e Verd, 2595AA, The Hague, Netherlands.

Declares that the MDD described hereafter

Product name and model:

ELFM3600 - Non-Woven Face Mask, Ear-loop, 3-ply, non-sterile

Meet the provisions of the Council Directive 93/42/EEC & 2007/47/CE which apply to them.

Harmonized standards or normative documents: EN 14683:2019+AC:2019 TYPE IIR BFE ≥98%

The medical device has been assigned to class I according to Annex VII of the Directive 2007/47/CE. It bears the mark:

CONFORMITY ASSESSMENT ROUTE: Annex VII

This Declaration of conformity is valid in connection with the release document for the respective batch of produced devices.

The above mentioned declaration of conformity is exclusively under the responsibility of: ELDO CARE CORPORATION LIMITED

Place , date Shanghai, China Sep. 9th, 2020 Legally binding signature, Function Co-founder & CEO

Ellen Wang

EC Declaration of Conformity

Authorized Signature(s)

CO., LIMITED

For and on behalf of ELDO CARE



ELDO CODE	GIMA CODE	PRODUCT DESCRIPTION	PACKING
ELFM3600	20670	GISAFE – color: light blue Surgical mask 3ply - Type IIR non-woven - earloop- non sterile – singly packed	box 50pcs
ELFM3600	20700	GISAFE – color: light blue Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	box 50pcs
ELFM3600	20695	GISAFE – color: light blue Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	flowpack 10pcs
ELFM3600	20697	GISAFE – color: pink Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	box 50pcs
ELFM3600	20694	GISAFE – color: pink Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	flowpack 10pcs
ELFM3600	20698	GISAFE – color: white Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	box 50pcs
ELFM3600	20699	GISAFE – color: light green Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	box 50pcs
ELFM3600	20685	GISAFE – fantasy: cartoon Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	flowpack 10pcs
ELFM3600	20686	GISAFE – fantasy: peace Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	flowpack 10pcs
ELFM3600	20687	GISAFE – fantasy: stars Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	flowpack 10pcs
ELFM3600	20680	GISAFE – color: light blue Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	flowpack 10pcs
ELFM3600	20682	GISAFE – fantasy: cartoon Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	flowpack 10pcs
ELFM3600	20696	GISAFE – color: light blue Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	box 2000pcs (200 flowpack of 10 pcs)
ELFM3600	20701	GISAFE – color: black Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	box 50pcs
ELFM3600	20688	GISAFE - fantasy: colours Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	flowpack 10pcs
ELFM3600	20689	GISAFE – fantasy: macarones Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	flowpack 10pcs
ELFM3600	20690	GISAFE – fantasy: skulls Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	flowpack 10pcs
ELFM3600	20691	GISAFE – fantasy: military Surgical mask 3ply - Type IIR non-woven - earloop- non sterile	flowpack 10pcs

Page 2/2



GISAFE MASCHERINA CHIRURGICA FILTRANTE 98% 3 VELI tipo IIR con elastici - adulti - colors - flowpack

Codice: 20688

Categoria: Mascherine

Unità di vendita: conf. 10 pz.

Quantitativi minimi: 1

Dispositivo: Dispositivo medico

Classe: I

NSIS: 2043900

CND: T020601

EAN13: 8023279206883

Descrizione: MASCHERINE FILTRANTI 98% A 3 VELI - TIPO IIR con elastici

• Modello per adulti • Flowpack da 10

• Fantasia: colours

Mascherine chirurgiche leggere e confortevoli adatte per ospedali, studi dentistici e ambulatori

- vasta gamma: mascherine adulti in 5 colori o 7 fantasie; mascherine pediatriche in 1 colore o 1 fantasia

- disponibili in scatola da 50 o flowpack da 10

- nasello modellabile da adattare alla forma anatomica del viso

- senza fibra di vetro, efficienza di filtrazione BFE =98%

materiale non tessuto, senza lattice
1° velo esterno 28 gsm non tessuto

- 2° velo 25 gsm filtro meltblown

- 3° velo interno 25 gsm non tessuto

 $Scatola\ multilingue:\ GB,\ FR,\ IT,\ ES,\ PT,\ DE,\ SE,\ GR,\ RO,\ Arabo.$

Caratteristiche tecniche: BFE% = 98

Pressione differenziale (Pa/cm2) < 60

CfU/g < 30

Dimensioni: 17,5 x 9,5 cm (adulti)

Conforme a EN 14683:2019 + AC:2019; DDM 93/42/CEE Classe I







国际互认 **TESTING** CNAS L0599

Date: January 18,2021

Page 1 of 3

ELDO CARE CORPORATION LIMITED

RM.1902, EASEY COMM. BLDG., 253-261 HENNESSY ROAD, WANCHAI, HONGKONG

The following sample(s) was/were submitted and identified on behalf of the client as:

Sample Description (A)GIMA BRANDED 20688 - 3 PLY SURGICAL MASK DISPOSABLE-

EARLOOP -FANTASY: COLOURS

Composition (A)67% non-woven, 33% melt-blown

Sample Color (A)FANTASY: COLOURS

Style No. ELFM3600 Model No. **ELFM3600** Lot No. Not provided

Manufacturer **ELDO CARE CORPORATION LIMITED**

Country of Origin China Type/ Level IIR

Sample Dimension 17.5X9.5CM

Test Performed Selected test(s) as requested by applicant

Sample Receiving Date Dec 28, 2020

Testing Period Dec 28, 2020 - Jan 18, 2021

Unless otherwise stated the results shown in this test report refer only to the Test Result(s)

sample(s) tested, for further details, please refer to the following page(s).

Signed for and on behalf of

SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd Testing Center

Helen Xuan (Authorized Signatory)





SL52045325756301TX

Date: January 18,2021

Page 2 of 3

Test Result

EN 14683:2019+AC:2019 Medical Face Masks-Requirements and Test Methods

Clause 5.2 Performance Requirement

Clause 5.2.2 Bacterial Filtration Efficiency (BFE)

(EN 14683:2019+AC:2019 Annex B)

Sample: A

Test Side : Inside

Test Area : Approximately 60 cm²

Flow Rate : 28.3 L/min

Pre-Conditioning : Minimum of 4 hours at 21±5°C and 85±5% R.H.

Dimensions of test specimen : ~178mm x 155mm

Positive Control Average : 2827 CFU
Negative Monitor Count : < 1 CFU
Mean Particle Size : 3.0 ±0.3µm

Test bacteria : Staphylococcus aureus ATCC 6538

Test Item	Specimen No.	Result
	1	99.9%
Postorial Filtration Efficiency	2	99.9%
Bacterial Filtration Efficiency (BFE)	3	99.9%
(BFC)	4	99.9%
	5	99.9%

Remark:

- 1) Performance Requirement: Type I≥95%, Type II≥98%, Type IIR≥98%
- 2) The number of specimens that shall be tested is minimum 5, but can be greater and shall be increased if necessary to allow for an AQL(Acceptable Quality Level) of 4%.



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/en/Terms-and-Conditions.aspx and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <a href="http://www.sgs.com/en/Terms-and-Conditions/Terms-and-Cond

3rd Building, No.889, Yishan Road, Xuhui District Shanghai, China 200233

中国·上海·徐汇区宜山路889号3号楼 邮编: 200233 t (86-21) 61402666

t (86-21) 61402666 f (86-21) 64958763 t (86-21) 61402666 f (86-21) 64958763



SL52045325756301TX

Date: January 18,2021

Page 3 of 3







The statement of conformity in this test report is only based on measured values by the laboratory and does not take their uncertainties into consideration.

End of Report



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/en/Terms-and-Conditions.aspx and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <a href="http://www.sgs.com/en/Terms-and-Conditions/Terms-and-Documents-Terms-and-Conditions/Terms-and-Documents-Terms-and-Conditions/Terms-and-Documents-Terms-and-Conditions/Terms-and-Condit

3rdBuilding,No.889,Yishan Road,Xuhui District Shanghai,China 200233 中国・上海・徐汇区宜山路889号3号楼 邮编: 200233 t (86-21) 61402666 t (86-21) 61402666

f (86–21) 64958763 f (86–21) 64958763





Date: January 13,2021



中国认可 国际互认 检测 TESTING CNAS L0599

Page 1 of 3

Test Report

SL52045325758901TX

ELDO CARE CORPORATION LIMITED

RM.1902, EASEY COMM. BLDG., 253-261 HENNESSY ROAD, WANCHAI, HONGKONG

The following sample(s) was/were submitted and identified on behalf of the client as:

Sample Description : (A)GIMA BRANDED 20688 - 3 PLY SURGICAL MASK DISPOSABLE-

EARLOOP -FANTASY: COLOURS

Composition : (A)67% non-woven, 33% melt-blown

Sample Color : (A)FANTASY: COLOURS

Style No. : ELFM3600

Model No. : ELFM3600

Lot No. : Not provided

Manufacturer : ELDO CARE CORPORATION LIMITED

Country of Origin : China Type/ Level : IIR

Sample Dimension : 17.5X9.5CM

Test Performed : Selected test(s) as requested by applicant

Sample Receiving Date : Dec 28, 2020


Testing Period : Dec 28, 2020 - Jan 13, 2021

Test Result(s) : Unless otherwise stated the results shown in this test report refer only to the

sample(s) tested, for further details, please refer to the following page(s).

Signed for and on behalf of

SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd Testing Center







SL52045325758901TX

Date: January 13,2021

Page 2 of 3

Test Result

EN 14683:2019+AC:2019 Medical Face Masks-Requirements and Test Methods

Clause 5.2 Performance Requirement

Clause 5.2.3 Breathability

(EN 14683 :2019+AC:2019 Annex C)

Sample: A

Test Side Randomly test in different location (1 around and 4 away from the centric

point) on each of the 5 masks

Pre-Conditioning Minimum of 4 hours at 21±5°C and 85±5% R.H.

Test Area 4.9 cm² Flow Rate 8 I/min

Specimen No.	Test Area No.	Different Pressure for each tested area (Pa/cm²)	The average value for each test specimen (Pa/cm²)
	1-1	44.4	
	1-2	43.9	
1	1-3	44.9	43
	1-4	38.5	
	1-5	41.7	
	2-1	40.5	
	2-2	41.9	
2	2-3	43.9	45
	2-4	51.7	
	2-5	44.8	
3	3-1	56.5	
	3-2	51.3	
	3-3	46.7	48
	3-4	40.3	
	3-5	46.1	
4	4-1	47.9	
	4-2	48.5	
	4-3	50.3	47
	4-4	40.7	
	4-5	47.8	
5	5-1	47.1	
	5-2	42.1	
	5-3	53.3	48
	5-4	50.8	
	5-5	44.2	

Remark:

- 1) Performance Requirement: Type I<40 Pa/cm², Type II<40 Pa/cm², Type IIR<60 Pa/cm²
- The number of specimens that shall be tested is minimum 5, but can be greater and shall be increased if necessary to allow for an AQL(Acceptable Quality Level) of 4%.



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/en/Terms-and-Conditions.aspx, and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at http://www.sgs.com/en/Terms-and-Conditions/Terms-and-Document.aspx. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@cgs.com

3rdBuilding, No. 889, Yishan Road, Xuhui District Shanghai, China 200233 中国・上海・徐汇区宜山路889号3号楼 邮编: 200233 t (86-21) 61402666 t (86-21) 61402666

f (86-21) 64958763

e sgs.china@sgs.com



SL52045325758901TX

Date: January 13,2021

Page 3 of 3





The statement of conformity in this test report is only based on measured values by the laboratory and does not take their uncertainties into consideration.

End of Report



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/en/Terms-and-Conditions.apx, and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <a href="http://www.sgs.com/en/Terms-and-Conditions/Terms-and-Condi









SL52045325754801TX

Date: January 13,2021

Page 1 of 3

ELDO CARE CORPORATION LIMITED

RM.1902, EASEY COMM. BLDG., 253-261 HENNESSY ROAD, WANCHAI, HONGKONG

The following sample(s) was/were submitted and identified on behalf of the client as:

Sample Description : (A)GIMA BRANDED 20688 - 3 PLY SURGICAL MASK DISPOSABLE-

EARLOOP -FANTASY: COLOURS

Composition : (A)67% non-woven, 33% melt-blown

Sample Color : (A)FANTASY: COLOURS

Style No. : ELFM3600

Model No. : ELFM3600

Lot No. : Not provided

Manufacturer : ELDO CARE CORPORATION LIMITED

Country of Origin : China Type/ Level : IIR

Sample Dimension : 17.5X9.5CM

Test Performed : Selected test(s) as requested by applicant

Sample Receiving Date : Dec 28, 2020


Testing Period : Dec 28, 2020 - Jan 13, 2021

Test Result(s) : Unless otherwise stated the results shown in this test report refer only to the

sample(s) tested, for further details, please refer to the following page(s).

Signed for and on behalf of

SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd Testing Center





Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/en/Terms-and-Conditions.apx, and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at http://www.sgs.com/en/Terms-and-Conditions/Terms-and-Document.aspx. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or, email: CN.Doccheck@cgs.com



SL52045325754801TX

Date: January 13,2021

Page 2 of 3

Test Result

EN 14683:2019+AC:2019 Medical Face Masks-Requirements and Test Methods

Clause 5.2 Performance Requirement

Clause 5.2.4 Splash Resistance

(ISO 22609:2004)

Sample: A

Test Blood Pressure : 16.0kPa

Pre-Conditioning : Minimum of 4 hours at 21±5°C and 85±5% R.H.

Distance of the mask to the tip of cannula : 300±10mm

Test Specimen#	Penetration on inside surface	Conclusion	Test Specimen#	Penetration on inside surface	Conclusion
1	None Seen	Pass	17	None Seen	Pass
2	None Seen	Pass	18	None Seen	Pass
3	None Seen	Pass	19	None Seen	Pass
4	None Seen	Pass	20	None Seen	Pass
5	None Seen	Pass	21	None Seen	Pass
6	None Seen	Pass	22	None Seen	Pass
7	None Seen	Pass	23	None Seen	Pass
8	None Seen	Pass	24	None Seen	Pass
9	None Seen	Pass	25	None Seen	Pass
10	None Seen	Pass	26	None Seen	Pass
11	None Seen	Pass	27	None Seen	Pass
12	None Seen	Pass	28	None Seen	Pass
13	None Seen	Pass	29	None Seen	Pass
14	None Seen	Pass	30	None Seen	Pass
15	None Seen	Pass	31	None Seen	Pass
16	None Seen	Pass	32	None Seen	Pass
Number of Pass:		32			
Overall result:		Acceptable			

Remark:

- 1) Performance Requirement Type I: N/A, Type II: N/A, Type IIR: ≥16.0kPa
- 2) Test was conducted within 60s after removal from conditioning chamber.
- 3) An acceptable quality limit of 4.0% is met for a single sampling plan when 29 or more of the 32 tested specimens show pass results.



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/en/Terms-and-Conditions.apx, and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <a href="http://www.sgs.com/en/Terms-and-Conditions/Terms-and-Condi

3⁴⁸Building,No.889,Yishan Road,Xuhui District Shanghai,China 200233 中国・上海・徐汇区宜山路889号3号楼 邮编: 200233

t (86-21) 61402666

f (86-21) 64958763



SL52045325754801TX

Date: January 13,2021

Page 3 of 3





The statement of conformity in this test report is only based on measured values by the laboratory and does not take their uncertainties into consideration.

End of Report



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/en/Terms-and-Conditions.apx, and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <a href="http://www.sgs.com/en/Terms-and-Conditions/Terms-and-Condi







Date:January 18,2021

Page 1 of 2

ELDO CARE CORPORATION LIMITED

RM.1902, EASEY COMM. BLDG., 253-261 HENNESSY ROAD, WANCHAI, HONGKONG

The following sample(s) was/were submitted and identified on behalf of the client as:

Sample Description : (A)GIMA BRANDED 20688 - 3 PLY SURGICAL MASK DISPOSABLE-

EARLOOP -FANTASY: COLOURS

Composition : (A)67% non-woven, 33% melt-blown

Sample Color : (A)FANTASY: COLOURS

Style No. : ELFM3600

Model No. : ELFM3600

Lot No. : Not provided

Manufacturer : ELDO CARE CORPORATION LIMITED

Country of Origin : China Type/ Level : IIR

Sample Dimension : 17.5X9.5CM

Test Performed : Selected test(s) as requested by applicant

Sample Receiving Date : Dec 28, 2020

Testing Period : Dec 28, 2020 - Jan 18, 2021

Test Result(s) : Unless otherwise stated the results shown in this test report refer only to the

sample(s) tested, for further details, please refer to the following page(s).

Signed for and on behalf of

SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd Testing Center

Sara Guo (Account Executive)

Dongjing Liu (Authorized Signatory)



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/en/Terms-and-Conditions.apx, and, for electronic format documents subject to Terms and Conditions for Electronic Documents at http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx, Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exoneriate to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com



SL52045325758001TX

Date:January 18,2021

Page 2 of 2

Test Result

EN 14683:2019+AC:2019 Medical Face Masks-Requirements and Test Methods

Clause 5.2 Performance Requirement

Clause 5.2.5 Microbial Cleanliness

(EN 14683:2019+AC:2019 Annex D and EN ISO 11737-1:2018)

Sample: A

Test Specimen#	Mask Weight(g)	Total Bioburden, (CFU/mask)	Total Bioburden, (CFU/g)
1#	3.60	51	14.17
2#	3.58	66	18.44
3#	3.62	69	19.06
4#	2.64	69	26.14
5#	3.65	60	16.44

Remark: Performance Requirement: Type I≤30 CFU/g, Type II≤30 CFU/g, Type IIR≤30 CFU/g







The statement of conformity in this test report is only based on measured values by the laboratory and does not take their uncertainties into consideration.

End of Report



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/en/Terms-and-Conditions.aspx and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <a href="http://www.sgs.com/en/Terms-and-Conditions/Terms-and-Cond

3^{rt}Building,No.889,Yishan Road,Xuhui District Shanghai,China 200233 中国・上海・徐江区宜山路889号3号楼 邮编: 200233 t (86-21) 61402666 f (86-21) 64958763 t (86-21) 61402666 f (86-21) 64958763